

PROFORMA REGARDING SAFE DRINKING WATER AND SANITARY CONDITION
CERTIFICATE.

No. DWSD-HB/2015-26/28

Date : 9/2/2026

It is certified that an inspection team headed by B.K. PRASAD, TECHNICAL
MANAGER, D.W. & SANITATION DIVISION, H2B (Name of Officers
with designation) from DISTRICT LEVEL WATER TESTING LAB (Name of
Department/Office) inspected the GALAXY HIGH SCHOOL,
PABRA ROAD, HAZARIBAGH (Name & Address of
the School) on 07/2/2026 and found that the GALAXY HIGH SCHOOL
(Name of school) has safe
drinking water facilities for the students and members of staff of the institution and is maintaining
the hygienic sanitation condition in the school building & the campus as per the norms
prescribed by the Central/State/U.T Govt.

The above valid for a period of 31/07/2026.

Signature with Seal : _____

B.K. Prasad
09/2/2026

Name

Quality Manager

District Level Water Testing Laboratory

Designation

D.W. & San. Div., Hazaribagh

To

GALAXY HIGH SCHOOL
PABRA ROAD, HAZARIBAGH
JHARKHAND-

(Name & Address of the Institution)



District Level Water Testing Laboratory, Hazaribagh
Office of the Executive, Engineer, Drinking Water and Sanitation Division, Hazaribagh
Contact no-7004844353. Email: dlwtl.hazaribagh@gmail.com
(Test Analysis Report)

Discipline - Chemical & Biological	Group - Water	Sub - Group - Drinking Water	
Issued To	Galaxy High School Pabra Road, Hazaribagh, Jharkhand-825319	SAMPLE CODE - 01	
		Report Issue Date : 09.02.2026	
Sample Description / Type	DRINKING WATER (G W)	Sample Collected by	LAB REP MAHESH KUMAR
Sample Location	FINAL OUTLET	Sample Quantity / Packing	1 L PVC Bottle
Date of Sample Received	07.02.2026	Sample Temperature	4 °C
Date of start of Analysis	07.02.2026	Humidity	50±15 % RH
Date of Completion Analysis :	09.02.2026	Sampling Procedure	N/A

Sl No.	Parameter	Unit	Result	Method Reference	Specification as per BIS - 10500 2012 (Second Revision) Amendment - June 2015	
					Acceptable Limit (Max)	Permissible Limit (In Absence of Alternate source) (Max)
(A) CHEMICAL PARAMETERS						
1	pH	-	7.40	APHA 24 th Edition 2023-4500-H ¹ B, (Electrometric Method)	6.5 - 8.5	No relaxation
2	Turbidity	NTU	2	APHA-24 th Edition 2023-2130 A, (Nephelometric method)	1	5
3	Chloride (as Cl)	mg/L Max	20	APHA-24 th Edition 2023-4500- B, (Argentometric method)	250	1000
4	Total Hardness as CaCO ₃	mg/L Max	230	APHA 24 th Edition 2023-2340 C (EDTA Titrimetric method)	200	600
5	Total Alkalinity as CaCO ₃	mg/L Max	200	APHA 24 th Edition 2023-2320 B, (Titration Method)	200	600
6	Calcium as Ca ⁺	mg/L Max	30.20	APHA 24 th Edition 2023-3500 Ca B, (EDTA Titrimetric Method)	75	200
7	Magnesium as Mg ⁺	mg/L Max	18.28	APHA 24 th Edition 2023-3500 Mg B, (EDTA Titrimetric Method)	30	100
8	Fluoride as F	mg/L Max	0.15	APHA 24 th Ed 2023-4500 F C, (Ion selective electrode method)	1.0	1.50
9	Colour	CU (Hazen)	2	APHA 24 th Edition 2023 (Visual Comparison Method)	5	15
10	Odour	-	Agreeable	IS 3025 (Part 5)	Agreeable	Agreeable
11	Iron as Fe	mg/L Max	0.23	APHA 24 th Edition 2023-4500 H ¹ B, (1,10 phenanthroline Method)	1.0	No relaxation
12	Nitrate as NO ₃	mg/L Max	BDL	APHA 24 th Edition 2023-4500 NO ₃ B, (Ultraviolet spectrophotometric screening Method)	45	No relaxation
13	Total Dissolved Solids at 180 °C	mg/L Max	150	APHA 24 th Edition 2023-2540 C (Total Dissolved Solids at 180 °C)	500	2000
(B) BIOLOGICAL PARAMETERS						
14	Total Coliform	CFU/100 mL	0	APHA 24 th Ed 2023-3500 Fe B (Membrane Filter Technique)	Shall not be detectable in any 100 mL Sample	
15	Thermotolerant (Fecal) Coliform/ E. Coli	CFU/100 mL	0	APHA 24 th Ed 2023-9222 D (Membrane Filter Technique)	Shall not be detectable in any 100 mL Sample	

Remarks:-On the basis of these tested parameters submitted/collected water samples confirms IS 10500 Yes / No ()

Specific contractual notes:-


- The results listed refer only to the tested sample and applicable parameter
- This report shall not be reproduced except in full without approval of the laboratory.
- This report results apply to the sample as received.
- This report, in full or in part, shall not be used for advertising or as evidence in any court of law
- Tested Sample shall be retained up to seven days from the date of issue of the test report
- The liability of the laboratory is limited to the invoiced amount
- All disputes are subjected to the Hazaribagh Jurisdiction.

Manager)

8 NOTE - BDL=Below Detection Limit

Reviewed by Technical Manager (B.K.Prasad)

End of Report.....


 09/2/2026
 Authorised Signatory
 (Abhishek Kumar, Quality
 Quality Manager
 District Level Water Testing Laboratory,
 D.W. & San. Div., Hazaribagh)